

B. Radioactive Material Sink Policy

1. NO radioactive material may be disposed through the sanitary sewer. Any radioactivity above background levels in any sink at NCI-Frederick is in direct violation of NRC license conditions.
2. Do not clean or rinse any contaminated equipment in the sink. Contaminated equipment must be rinsed off into an appropriate container and the waste water collected for disposal as aqueous radioactive waste, as instructed in [section IV: M](#) (Disposal of Radioactive Material) of this manual.